

Template for PCR pathogen Detection

Positive Control Template SE1

Code No. S035 **Size:** **5 ng**
Conc.: **100 pg/μl**

* 2 years from date of receipt under proper storage conditions.

Detectable Gene :

This product contains positive control template for PCR detection of pathogenic genes. For the detection of sea (*entA*) gene encoding staphylococcal enterotoxin A, use in conjunction with Primer Set SEA-1&2 (Cat. #S009), and for the detection of seb (*entN*) gene encoding staphylococcal enterotoxin B, use in conjunction with Primer Set SEB-1&2 (Cat. #S010).

Amplified Fragment :

with Primer Set SEA (Cat. #S009) : 695 bp
with Primer Set SEB (Cat. #S010) : 694 bp

Form : Supplied in TE buffer.

10 mM Tris-HCl, pH 8.0
1 mM EDTA

Storage: -20°C

Application Example :

1. Reaction mixture for PCR (total 50 μl)

<i>TaKaRa Taq</i> [™] (5 U/μl)	0.25 μl
10X PCR Buffer* ¹	5 μl
dNTP Mixture* ² (2.5 mM each)	4 μl
Primer 1	0.5 μl
Primer 2	0.5 μl
Positive Control Template	0.5 μl
Sterile purified water	39.25 μl
Total	50 μl

(Reaction mixture shall be prepared on ice.)

*1 Supplied with Primer Set.

*2 Supplied with *TaKaRa Taq* (Cat. #R001A).

2. PCR Condition

94°C	1 min	} 35 cycles
55°C	1 min	
72°C	1 min	

TaKaRa Taq is a trademark of Takara Bio Inc.

Note

This product is for research use only. It is not intended for use in therapeutic or diagnostic procedures for humans or animals. Also, do not use this product as food, cosmetic, or household item, etc. Takara products may not be resold or transferred, modified for resale or transfer, or used to manufacture commercial products without written approval from Takara Bio Inc.

If you require licenses for other use, please contact us by phone at +81 77 565 6973 or from our website at www.takara-bio.com.

Your use of this product is also subject to compliance with any applicable licensing requirements described on the product web page. It is your responsibility to review, understand and adhere to any restrictions imposed by such statements.

All trademarks are the property of their respective owners. Certain trademarks may not be registered in all jurisdictions.

Template for PCR pathogen Detection

Positive Control Template SE1

Code No. S035 容量： 5 ng
濃度： 100 pg/ μ l

※ 適切に保存し、受取り後 2 年を目途にご使用ください。

● 製品説明

特殊細菌検出用 Primer Set を用いて病原菌検出 PCR を行う際に、PCR 反応が正常に行われているかを確認するためのポジティブコントロールテンプレートである。(PCR 産物は、病原菌由来の増幅産物とサイズが異なる。)

本製品は、黄色ブドウ球菌エンテロトキシン A 遺伝子検出用 Primer Set SEA-1&2 (製品コード S009)、もしくは、黄色ブドウ球菌エンテロトキシン B 遺伝子検出用 Primer Set SEB-1&2 (製品コード S010) のテンプレートとして使用する。

● 増幅産物

PrimerSet SEA (製品コード S009) のテンプレートとして用いた場合：695 bp
PrimerSet SEB (製品コード S010) のテンプレートとして用いた場合：694 bp

● 形状

10 mM Tris-HCl, pH8.0
1 mM EDTA

● 保存

- 20°C

● 使用例

1. 反応液組成 (total 50 μ l)

TaKaRa Taq (5 U/ μ l)	0.25 μ l
10 × PCR Buffer*1	5 μ l
dNTP Mixture*2 (2.5 mM each)	4 μ l
Primer 1	0.5 μ l
Primer 2	0.5 μ l
Positive Control Template	0.5 μ l
滅菌精製水	39.25 μ l
Total	50 μ l

(反応液は氷冷下で調製する)

* 1：特殊細菌検出用 Primer Set に添付している。

* 2：TaKaRa Taq (製品コード R001A) に添付している。

2. PCR 条件

94°C	1 min.	} 35 cycles
55°C	1 min.	
72°C	1 min.	

● 注意

本製品は研究用として販売しております。ヒト、動物への医療、臨床診断用には使用しないようご注意ください。また、食品、化粧品、家庭用品等として使用しないでください。
タカラバイオの承認を得ずに製品の再販・譲渡、再販・譲渡のための改変、商用製品の製造に使用することは禁止されています。
ライセンスに関する情報は弊社ウェブカタログをご覧ください。
本データシートに記載されている会社名および商品名などは、各社の商号、または登録済みもしくは未登録の商標であり、これらは各所有者に帰属します。

v201812Da